



Sample Details

Sample Name: 1
SOP Name: mansettings.nano
General Notes:

File Name: controllo ravazzolo gero... Dispersant Name: Water
Record Number: 11 Dispersant RI: 1,330
Material RI: 1,46 Viscosity (cP): 0,8872
Material Absorbtion: 3,320 Measurement Date and Time: giovedì 7 novembre 2024 1...

System

Temperature (°C): 25,0 Duration Used (s): 60
Count Rate (kcps): 272,6 Measurement Position (mm): 3,00
Cell Description: Disposable micro cuvette (40... Attenuator: 6

Results

	Size (d.n...	% Intensity:	St Dev (d.n...
Z-Average (d.nm): 217,0	Peak 1: 237,5	93,0	73,86
Pdl: 0,380	Peak 2: 49,83	7,0	8,915
Intercept: 0,793	Peak 3: 0,000	0,0	0,000

Result quality Good

