

Size Quality Report

v2.0



Malvern Instruments Ltd - © Copyright 2008

Sample Name: forniz controllo 1

SOP Name: mansettings.nano

File Name: ravazzolo.dts

Record Number: 3

Measurement Date and Time: giovedì 7 novembre 2024 12:08:23

Temperature (°C): 24,9

RESULT DOES NOT MEET QUALITY CRITERIA

Sample too polydisperse for cumulant analysis - suggest rely on distribution analysis

Cumulant fit error high

- * Data quality too poor for cumulant analysis
- * Sample too polydisperse for cumulant analysis

