

Size Quality Report

v2.0



Malvern Instruments Ltd - © Copyright 2008

Sample Name: cebin controllo 1

SOP Name: mansettings.nano

File Name: ravazzolo.dts

Record Number: 7

Measurement Date and Time: giovedì 7 novembre 2024 12:22:38

Temperature (°C): 25,0

RESULT DOES NOT MEET QUALITY CRITERIA

Multimodal fit error high

- * Data quality too poor for distribution analysis
- * Sample too polydisperse for distribution analysis

Cumulant fit error high

- * Data quality too poor for cumulant analysis
- * Sample too polydisperse for cumulant analysis

In range figure is low (85%)

- * Presence of large or sedimenting particles
- * Sample fluorescence
- * Sample absorbance (coloured samples)

