

# Size Quality Report

v2.0



Malvern Instruments Ltd - © Copyright 2008

**Sample Name:** AHST0057\_0.1mg/mL\_H2OmQ

**SOP Name:** mansettings.nano

**File Name:** ripetute cebin.dts

**Record Number:** 4

**Measurement Date and Time:** giovedì 14 novembre 2024 11:43:24

**Temperature (°C):** 25,0

## RESULT DOES NOT MEET QUALITY CRITERIA

Sample too polydisperse for cumulant analysis - suggest rely on distribution analysis

Cumulant fit error high

- \* Data quality too poor for cumulant analysis
- \* Sample too polydisperse for cumulant analysis

